

BEST AVAILABLE COPY**AMENDMENTS TO THE CLAIMS**

Please amend claims 1, 21, 41, 64, and 84, and cancel claim 55-63. This listing of claims will replace all prior versions, and listings, of claims in the application.

What is claimed is:

- 1 1. (Currently Amended) A monitor that can detect a plurality of trace molecules,
2 comprising:
3 a housing with an ionizing chamber that is approximately at one atmosphere and a single
4 sample inlet that allows a sample to flow into said ionizing chamber;
5 a photoionizer that is coupled to said ionizing chamber and can be activated and
6 deactivated to ionize the sample;
7 an electrospray ionizer coupled to said ionizing chamber and can be activated and
8 deactivated to ionize the sample;
9 a switch that ~~controls the operation~~ activates and deactivates of said photoionizer and said
10 electrospray ionizer to control different modes of operation; and,
11 a detector that is coupled to said ionizing chamber.

- 1 2. (Original) The monitor of claim 1, wherein said electrospray ionizer includes a
2 vaporizer.

- 1 3. (Original) The monitor of claim 1, further comprising a chemical ionizer coupled
2 to said ionizing chamber and said switch.

Attorney Docket No. 155692-0034(P002CZ)
App. No. 10/672,958
1460468

- 2 -

BJY:sml

1 4. (Original) The monitor of claim 3, wherein said chemical ionizer includes a
2 vaporizer.

1 5. (Original) The monitor of claim 2, further comprising a vacuum interface
2 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
3 that is orthogonal to said electrospray ionizer vaporizer.

1 6. (Original) The monitor of claim 4, further comprising a vacuum interface
2 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
3 that is orthogonal to said electrospray ionizer vaporizer.

1 7. (Original) The monitor of claim 1, further comprising a processor that controls
2 said switch.

3 8. (Original) The monitor of claim 1, wherein said switch operates in a mode where
4 said electrospray ionizer and said photoionizer are sequentially activated.

1 9. (Original) The monitor of claim 1, wherein said switch operates in a mode
2 where said electrospray ionizer and said photoionizer are simultaneously activated.

1 10. (Original) The monitor of claim 8, wherein said switch operates in a mode
2 wherein said electrospray ionizer and said photoionizer each generates a positive ion, then each
3 generates a negative ion.

1 11. (Original) The monitor of claim 8, wherein said switch operates in a mode
2 wherein said electrospray ionizer and said photoionizer each generates pairs of positive and
3 negative ions sequentially in time.

1 12. (Original) The monitor of claim 1, wherein said switch operates in a mode where
2 said photoionizer is on and said electrospray ionizer is switched between on and off states.

1 13. (Original) The monitor of claim 1, wherein said switch operates in a mode
2 wherein said electrospray ionizer is on and said photoionizer is switched between on and off
3 states.

1 14. (Original) The monitor of claim 1, wherein said electrospray ionizer and said
2 photoionizer each have an electrode that is supplied a voltage from a same voltage source.

1 15. (Original) The monitor of claim 9, further comprising a chemical ionizer that is
2 coupled to said switch and generates a positive ion sequentially with said electrospray ionizer
3 and said photoionizer, and then generates a negative ion sequentially with said electrospray
4 ionizer and said photoionizer.

1 16. (Original) The monitor of claim 10, further comprising a chemical ionizer that is
2 coupled to said switch and generates a positive and negative ion pair sequentially with said
3 electrospray ionizer and said photoionizer.

1 17. (Original) The monitor of claim 1, further comprising a valve that controls a
2 flow of a sample through an inlet of said electrospray ionizer and an inlet of said photoionizer.

1 18. (Original) The monitor of claim 17, wherein said valve sequentially allows the
2 sample to flow through said electrospray ionizer inlet and said photoionizer inlet.

1 19. (Original) The monitor of claim 17, wherein said valve simultaneously allows
2 the sample to flow through said electrospray ionizer inlet and said photoionizer inlet.

1 20. (Original) The monitor of claim 17, wherein said valve creates different flow
2 rates through said electrospray ionizer inlet and said photoionizer inlet.

1 21. (Currently Amended) A monitor that can detect a plurality of trace molecules,
2 comprising:
3 a housing with an ionizing chamber that is approximately at one atmosphere and a single
4 sample inlet that allows a sample to flow into said ionizing chamber;
5 a photoionizer that is coupled to said ionizing chamber and can be activated and
6 deactivated to ionize the sample;
7 an electrospray ionizer coupled to said ionizing chamber and can be activated and
8 deactivated to ionize the sample;
9 switch means for controlling the operation of said photoionizer and said electrospray
10 ionizer to control different modes of operation; and,
11 a detector that is coupled to said ionizing chamber.

1 22. (Original) The monitor of claim 21, wherein said electrospray ionizer includes a
2 vaporizer.

1 23. (Original) The monitor of claim 21, further comprising a chemical ionizer
2 coupled to said ionizing chamber and said switch means.

1 24. (Original) The monitor of claim 23, wherein said chemical ionizer includes a
2 vaporizer.

1 25. (Original) The monitor of claim 22, further comprising a vacuum interface
2 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
3 that is orthogonal to said electrospray ionizer vaporizer.

1 26. (Original) The monitor of claim 24, further comprising a vacuum interface
2 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
3 that is orthogonal relative to said electrospray ionizer vaporizer.

1 27. (Original) The monitor of claim 21, further comprising a processor that controls
2 said switch means.

3 28. (Original) The monitor of claim 21, wherein said switch means operates in a
4 mode where said electrospray ionizer and said photoionizer are sequentially activated.

1 29. (Original) The monitor of claim 21, said switch means operates in a mode where
2 said electrospray ionizer and said photoionizer are simultaneously activated.

1 30. (Original) The monitor of claim 28, wherein said switch means operates in a
2 mode wherein said electrospray ionizer and said photoionizer each generates a positive ion, then
3 each generates a negative ion.

1 31. (Original) The monitor of claim 28, wherein said switch means operates in a
2 mode wherein said electrospray ionizer and said photoionizer each generates pairs of positive
3 and negative ions sequentially in time.

1 32. (Original) The monitor of claim 21, wherein said switch means operates in a
2 mode where said photoionizer is on and said electrospray ionizer is switched between on and off
3 states.

1 33. (Original) The monitor of claim 21, wherein said switch means operates in a
2 mode wherein electrospray ionizer is on and said photoionizer is switched between on and off
3 states.

1 34. (Original) The monitor of claim 21, wherein said electrospray ionizer and said
2 photoionizer each have an electrode that is supplied a voltage from a same voltage source.

1 35. (Original) The monitor of claim 30, further comprising a chemical ionizer that is
2 coupled to said switch means to generate a positive ion sequentially with said electrospray
3 ionizer and said photoionizer, and then generates a negative ion sequentially with said
4 electrospray ionizer and said photoionizer.

1 36. (Original) The monitor of claim 30, further comprising a chemical ionizer that is
2 coupled to said switch means to generate a positive and negative pair of ions sequentially with
3 said electrospray ionizer and said photoionizer.

1 37. (Original) The monitor of claim 21, further comprising a valve that controls a
2 flow of a sample through an inlet of said electrospray ionizer and an inlet of said photoionizer.

1 38. (Original) The monitor of claim 37, wherein said valve sequentially allows the
2 sample to flow through said electrospray ionizer inlet and said photoionizer inlet.

1 39. (Original) The monitor of claim 37, wherein said valve simultaneously allows
2 the sample to flow through said electrospray ionizer inlet and said photoionizer inlet.

1 40. (Original) The monitor of claim 37, wherein said valve creates different
2 flowrates through said electrospray ionizer inlet and said photoionizer inlet.

1 41. (Currently Amended) A method for detecting a plurality of trace molecules,
2 comprising:

3 introducing a sample into an ionizing chamber through a single sample inlet;
4 ionizing a trace molecule within the sample with a photoionizer at approximately
5 atmospheric pressure;
6 ionizing a trace molecule within the sample with an electrospray ionizer at approximately
7 atmospheric pressure;
8 detecting the ionized trace molecules; and,
9 switching a mode of operation of the photoionizer and the electrospray ionizer by
10 deactivating the photoionizer or the electrospray ionizer.

1 42. (Original) The method of claim 41, further comprising vaporizing a sample that
2 contains the trace molecules.

1 43. (Original) The method of claim 41, further comprising ionizing a trace molecule
2 with a chemical ionizer at approximately atmospheric pressure.

1 44. (Original) The method of claim 41, wherein the mode includes activating the
2 electrospray ionizer and the photoionizer sequentially.

1 45. (Original) The method of claim 41, wherein the mode includes activating the
2 electrospray ionizer and the photoionizer simultaneously.

1 46. (Original) The method of claim 44, wherein the mode includes activating the
2 electrospray ionizer and the photoionizer so that each generates a positive ion, then each
3 generates a negative ion.

1 47. (Original) The method of claim 44, wherein the mode includes activating the
2 electrospray ionizer and the photoionizer so that each generates pairs of positive and negative
3 ions sequentially in time.

1 48. (Original) The method of claim 41, wherein the mode includes maintaining the
2 photoionizer on, while switching the electrospray ionizer between on and off states.

1 49. (Original) The method of claim 41, wherein the mode includes maintaining the
2 electrospray ionizer on, while switching the photoionizer between on and off states.

1 50. (Original) The method of claim 44, further comprising ionizing a trace molecule
2 with a chemical ionizer in a mode where the chemical ionizer generates a positive ion
3 sequentially with the electrospray ionizer and the photoionizer, and then generates a negative ion
4 sequentially with the electrospray ionizer and the photoionizer.

1 51. (Original) The method of claim 44, further comprising ionizing a trace molecule
2 with a chemical ionizer in a mode where the chemical ionizer generates a positive and negative
3 ion pair sequentially with the electrospray ionizer and photoionizer.

1 52. (Original) The method of claim 41, wherein a sample with the trace molecules
2 sequentially flows through an electrospray ionizer inlet and a photoionizer inlet.

1 53. (Original) The method of claim 41, wherein a sample with the trace molecules
2 simultaneously flows through an electrospray ionizer inlet and a photoionizer inlet.

1 54. (Original) The method of claim 41, wherein a sample with the trace molecules
2 flows through an electrospray ionizer inlet and a photoionizer inlet at different flow rates.

1 Claims 55-63 (Canceled)

1 64. (Currently Amended) A monitor that can detect a plurality of trace molecules,
2 comprising:
3 a housing with an ionizing chamber that is approximately at one atmosphere and a single
4 sample inlet that allows a sample to flow into said ionizing chamber;
5 a photoionizer that is coupled to said ionizing chamber and can be activated and
6 deactivate to ionize the sample;
7 a chemical ionizer coupled to said ionizing chamber and can be activated and deactivate
8 to ionize the sample;
9 a switch that controls the operation of said photoionizer and said chemical ionizer to
10 control different modes of operation; and,
11 a detector that is coupled to said ionizing chamber.

1 65. (Original) The monitor of claim 64, wherein said chemical ionizer includes a
2 vaporizer.

3 66. (Original) The monitor of claim 65, further comprising a vacuum interface
4 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
5 that is orthogonal to said chemical ionizer vaporizer.

1 67. (Original) The monitor of claim 64, further comprising a processor that controls
2 said switch.

3 68. (Original) The monitor of claim 64, wherein said switch operates in a mode
4 where said chemical ionizer and said photoionizer are sequentially activated.

1 69. (Original) The monitor of claim 64, wherein said switch operates in a mode
2 where said chemical ionizer and said photoionizer are simultaneously activated.

1 70. (Original) The monitor of claim 68, wherein said switch operates in a mode
2 wherein said chemical ionizer and said photoionizer each generates a positive ion, then each
3 generates a negative ion.

1 71. (Original) The monitor of claim 68, wherein said switch operates in a mode
2 wherein said chemical ionizer and said photoionizer each generates pairs of positive and negative
3 ions sequentially in time.

1 72. (Original) The monitor of claim 64, wherein said switch operates in a mode
2 where said photoionizer is on and said chemical ionizer is switched between on and off states.

1 73. (Original) The monitor of claim 64, wherein said switch operates in a mode
2 wherein said chemical ionizer is on and said photoionizer is switched between on and off states.

1 74. (Currently Amended) A monitor that can detect a plurality of trace molecules,
2 comprising:
3 a housing with an ionizing chamber that is approximately at one atmosphere and a single
4 sample inlet that allows a sample to flow into said ionizing chamber;
5 a photoionizer that is coupled to said ionizing chamber and can be activated and
6 deactivated to ionize the sample;
7 a chemical ionizer coupled to said ionizing chamber and can be activated and deactivated
8 to ionize the sample;
9 switch means for controlling the operation of said photoionizer and said chemical ionizer
10 to control different modes of operation; and,
11 a detector that is coupled to said ionizing chamber.

1 75. (Original) The monitor of claim 74, wherein said chemical ionizer includes a
2 vaporizer.

1 76. (Original) The monitor of claim 74, further comprising a vacuum interface
2 coupled to said ionizing chamber and said detector, said vacuum interface having an entrance
3 that is orthogonal to said chemical ionizer vaporizer.

1 77. (Original) The monitor of claim 74, further comprising a processor that controls
2 said switch means.

3 78. (Original) The monitor of claim 74, wherein said switch means operates in a
4 mode where said chemical ionizer and said photoionizer are sequentially activated.

1 79. (Original) The monitor of claim 74, said switch means operates in a mode where
2 said chemical ionizer and said photoionizer are simultaneously activated.

1 80. (Original) The monitor of claim 78, wherein said switch means operates in a
2 mode wherein said chemical ionizer and said photoionizer each generates a positive ion, then
3 each generates a negative ion.

1 81. (Original) The monitor of claim 78, wherein said switch means operates in a
2 mode wherein said chemical ionizer and said photoionizer each generates pairs of positive and
3 negative ions sequentially in time.

1 82. (Original) The monitor of claim 74, wherein said switch means operates in a
2 mode where said photoionizer is on and said chemical ionizer is switched between on and off
3 states.

1 83. (Original) The monitor of claim 74, wherein said switch means operates in a
2 mode wherein chemical ionizer is on and said photoionizer is switched between on and off states.

Attorney Docket No. 155692-0034(P002X2)
App. No. 10/672,958
1460468

1 84. (Currently Amended) A method for detecting a plurality of trace molecules,
2 comprising:
3 introducing a sample into an ionizing chamber through a single sample inlet;
4 ionizing a trace molecule within the sample with a photoionizer at approximately
5 atmospheric pressure;
6 ionizing a trace molecule with the same with an chemical ionizer at approximately
7 atmospheric pressure;
8 detecting the ionized trace molecules; and,
9 switching a mode of operation of the photoionizer and the chemical ionizer by
10 deactivating the photoionizer or the chemical ionizer.

1 85. (Original) The method of claim 84, further comprising vaporizing a sample that
2 contains the trace molecules.

1 86. (Original) The method of claim 84, wherein the mode includes activating the
2 chemical ionizer and the photoionizer sequentially.

1 87. (Original) The method of claim 84, wherein the mode includes activating the
2 chemical ionizer and the photoionizer simultaneously.

1 88. (Original) The method of claim 86, wherein the mode includes activating the
2 chemical ionizer and the photoionizer so that each generate a positive ion, then each generate a
3 negative ion.

Attorney Docket No. 155692-0034(P002X2)
App. No. 10/672,958
1460468

- 15 -

BJY:smj

1 89. (Original) The method of claim 86, wherein the mode includes activating the
2 chemical ionizer and the photoionizer so that each generate pairs of positive and negative ions
3 sequentially in time.

1 90. (Original) The method of claim 84, wherein the mode includes maintaining the
2 photoionizer on, while switching the chemical ionizer between on and off states.

1 91. (Original) The method of claim 84, wherein the mode includes maintaining the
2 chemical ionizer on, while switching the photoionizer between on and off states.

**This Page is Inserted by IFW Indexing and Scanning
Operations and is not part of the Official Record**

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

- ☐ BLACK BORDERS
- ☐ IMAGE CUT OFF AT TOP, BOTTOM OR SIDES
- ☐ FADED TEXT OR DRAWING
- ☐ BLURRED OR ILLEGIBLE TEXT OR DRAWING
- ☐ SKEWED/SLANTED IMAGES
- ☐ COLOR OR BLACK AND WHITE PHOTOGRAPHS
- ☐ GRAY SCALE DOCUMENTS
- ☒ LINES OR MARKS ON ORIGINAL DOCUMENT
- ☐ REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY
- ☐ OTHER: _____

IMAGES ARE BEST AVAILABLE COPY.

As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.